PrymaLab
PRYMALAB ANALYTICAL · IN-HOUSE LABORATORY
3705 Music Branch Rd, Ashland, KY 41102 · +1 (606) 465-9063 · prymalab.net

CERTIFICATE NO.
PL-26-00056
VERIFICATION KEY
5S96ABAD
Scan or enter both at prymalab.net/testing/verify

CERTIFICATE OF ANALYSIS · LOT RELEASE

Kisspeptin-10 10 mg

Lot **PL-2609-056** · Lyophilized powder · Label claim **10 mg/vial**

TESTED IN-HOUSE BY PRYMALAB
In-house analytical result, not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL-2609-056	2026-09-18	2026-09-19	2026-10-03
RETEST / EXPIRY	**UNITS IN LOT / SAMPLED**	**STORAGE**	**CERTIFICATE TYPE**
2028-09-18	250 / 2	−20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND	CAS NUMBER	MOLECULAR FORMULA	THEORETICAL MASS (AVG.)
Kisspeptin-10 KP-10; human metastin 45–54	374675-21-5	$C_{63}H_{83}N_{17}O_{14}$	1302.4 Da

SEQUENCE
H-YNWNSFGLRF-NH_2 (10 residues; C-terminal amide)

PURITY · RP-HPLC, AREA %
99.24%
Specification 99.0% • main peak at 5.18 min

IDENTITY · OBSERVED MASS
1302.6 Da
Theoretical 1302.4 Da • +0.2 Da (153 ppm)

OVERALL CONCLUSION
PASS
Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Appearance	Visual inspection	White to off-white lyophilized powder	White lyophilized powder	PASS
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da	1302.6 Da	PASS
Purity	RP-HPLC, UV 214 nm, area%	99.0%	99.24%	PASS
Largest impurity	RP-HPLC, area%	1.0%	0.31% at 5.94 min	PASS
Water content	Karl Fischer	5.0%	1.8%	PASS
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.04 EU/mg	PASS
Sterility / bioburden	USP <71>, membrane filtration	No growth after 14 days	No growth; conforms	PASS

1 HPLC CHROMATOGRAM

Data file: KSP_PL-2609-056_20260919_01 · Column C18 150 × 4.6 mm · UV 214 nm



Baseline uncropped; main peak and impurities labeled with retention times. Peak table on page 2.

For laboratory research use only. Not for human or veterinary use.

2 HPLC METHOD · KSP_PL-2609-056_20260919_01

INSTRUMENT
Waters Alliance e2695, SN E2695B742
DATA SYSTEM
Empower 3, Build 3471
DETECTOR
Waters 2998 PDA
DETECTION WAVELENGTH
214 nm
COLUMN
C18, 150 × 4.6 mm, 3.5 µm
COLUMN PART / SERIAL
186003326 / USHX18211
MOBILE PHASE A
0.1% TFA in water
MOBILE PHASE B
0.1% TFA in MeCN
FLOW RATE
1.0 mL/min
COLUMN TEMPERATURE
30 °C
INJECTION VOLUME
20 µL
RUN TIME
10 min
SAMPLE DILUENT
Water, 0.1% TFA
SAMPLE CONCENTRATION
0.5 mg/mL
BLANK INJECTED
Yes
REFERENCE STANDARD
In-house Kisspeptin-10 reference, lot PL-RS-KSP-01, RT 5.16 min

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	95	5
3.0	95	5
7.0	55	45
8.0	20	80
9.0	20	80
10.0	95	5

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1	4.38	7,310	0.17
2	**5.18**	**4,267,320**	**99.24**
3	5.94	13,330	0.31
4	6.72	12,040	0.28
Total		4,300,000	100.00

Threshold 0.05% area · Blank subtracted: Y · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT
Waters Acquity SQD
IONIZATION / MODE
ESI, positive
OBSERVED M/Z (CHARGE)
1303.3 (1+ [M+H]) · 1325.3 (1+ [M+Na])
DECONVOLUTED MASS
1302.6 Da
THEORETICAL MASS
1302.4 Da
DIFFERENCE
+0.2 Da (153 ppm)
DATA FILE
KSP_PL-2609-056_MS_20260919_01
ACCEPTANCE
Within ± 1.0 Da

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content	PrymaLab	2026-09-20	RPT-KF-PL-2609-056
Endotoxin	PrymaLab	2026-09-21	RPT-ET-PL-2609-056
Sterility	PrymaLab	2026-10-02	RPT-ST-PL-2609-056
Peptide content	PrymaLab	2026-09-20	RPT-AA-PL-2609-056

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds
Analyst · performed the testing
Date: 2026-09-19

S. Calloway
Reviewer · checked against raw data
Date: 2026-10-02

M. Phelps
Lab Director · approved for release
Date: 2026-10-03

Raw data retained at: Empower 3 (Build 3471), project PL-2609, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

This certificate applies only to the lot named above and describes the units sampled from it; it is not a guarantee for any other lot. Results were obtained by PrymaLab LLC's own laboratory (PrymaLab Analytical) and are not an independent third-party test unless an outside report is attached. Reference values (CAS, formula, theoretical mass, sequence) are literature values for the named compound. Authenticity: scan the QR code or enter certificate number PL-26-00056 and key 5S96ABAD at prymalab.net/testing/verify; any copy that does not match the record shown there is not a PrymaLab certificate. PrymaLab LLC products are supplied for laboratory research use only and not for human or veterinary use.